

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 07D2093663 | (X3) Date Survey Completed 10/07/2019 |
| Name of Provider or Supplier Precision Toxicology, Llc | Street Address, City, State 836 Farmington Ave Ste 212, West Hartford, CT | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D5209 | <p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to have a policy in place to assess the competency of all laboratory personnel. Findings include: 1. Review of the laboratory's competency records on 10/7/19 revealed the following: a. The laboratory did not have policy in place to assess the competency of the technical consultant (TC) and clinical consultant. b. Competency for the above laboratory personnel was not performed. 2. Staff interview with the TC on 10/7/19 at 10:10 AM confirmed the laboratory did not have a policy in place to assess the competency of the above laboratory personnel and they were not assessed.</p> |