

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 07D2102440	(X3) Date Survey Completed 03/09/2018
Name of Provider or Supplier Afc Urgent Care	Street Address, City, State 1918 Black Rock Turnpike, Fairfield, CT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	
D2000	ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801 Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994. This CONDITION is not met as evidenced by: No deficiency details available.