

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 08D0205587	(X3) Date Survey Completed 12/08/2021
Name of Provider or Supplier Panzer Dermatology & Associates	Street Address, City, State 537 Stanton-Christiana Road, Newark, DE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, inventory records review and staff interview the laboratory failed to discard current in use lot of expired Xylene reagents. Findings include: 1. During a tour of the Panzer Dermatology & Associates, PA testing area on December 8, 2021 at approximately 8:30 am, the tour revealed 4 of 4 current lot at the time of the survey of expired bottles of Xylene. a. Lot #: 087751; Expiration date: 9/30/21 2. An inventory document (Monthly Receipt Expense Log) detailed: a. Reagent Name: Xylene; Quantity: 4/Case Lot#: 087751; Expiration date: 9/30/21; Received date: 10/16/19 3. During the interview on December 8, 2021 at approximately 9:00 AM, the Testing Personnel #1, confirmed the above finding.</p>