

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 08D0903231	(X3) Date Survey Completed 08/01/2025
Name of Provider or Supplier Delaware Pediatrics Pa	Street Address, City, State 1403 Foulk Road, Suite 103, Wilmington, DE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	D0000 A Recertification Survey was conducted on August 1, 2025 at approximately 9:00AM. The laboratory was surveyed according to 42 CFR Part 493 Clinical Laboratory Improvement Amendments (CLIA) requirements. Deficiencies were identified as follows:
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on facility document review and interview, the laboratory failed to document the review and evaluation of results of proficiency testing (PT) for Strep Group A Identification. This was noted for 7 out of 7 (WSLH 2025 Bacti Viral 1, WSLH 2024 Bacti Viral 1, WSLH 2024 Bacti Viral 2, WSLH 2024 Bacti Viral 3, WSLH 2023 Bacti Viral 1, WSLH 2023 Bacti Viral 2, WSLH 2023, Bacti Viral 3) events reviewed. Findings included: Review of PT testing results from Wisconsin State Laboratory Hygiene (WSLH) PT Program revealed the Lab Director (LD) or designee had not signed as having reviewed the results for the following events: WSLH 2025 Bacti Viral 1, WSLH 2024 Bacti Viral 1, WSLH 2024 Bacti Viral 2, WSLH 2024 Bacti Viral 3, WSLH 2023 Bacti Viral 1, WSLH 2023 Bacti Viral 2, WSLH 2023, Bacti Viral 3. During an interview on 08/01/2025 at 10:40 AM, Head Nurse (HN) #4 was asked for documentation of the LD's review of the PT results and could not provide any documentation.</p>