

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 08D2212358	(X3) Date Survey Completed 02/07/2023
Name of Provider or Supplier Dermatology Partners - N Wilmington	Street Address, City, State 2000 Foulk Road, Suite A, Wilmington, DE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An initial survey was conducted on February 7, 2023 at approximately 09:15 am at Dermatology Partners - N. Wilmington. The laboratory was surveyed according to 42 CFR part 493 Clinical Laboratory Improvement Amendments (CLIA) requirements. Specific deficiencies are as follows:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: documentation and interview, the laboratory failed to perform Proficiency Testing (PTs) as required. Findings include: 1. At approximately 9:25 am on February 7, 2023 during document review, there was no second Proficiency Testing performed by the Testing Personnel (TP) in 2022. A PT was performed by the TP in January 2022, but none was performed or documented for the remainder of 2022 as required. 2. During the interview with the Laboratory Manager (LM), the LM confirmed that no second PT for 2022 was documented for the TP. 3. At the end of the survey at approximately 10:20 am no documentation could be provided as required for a second PT in 2022 by the TP.</p>