

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0270011	<b>(X3) Date Survey Completed</b>  02/02/2024
<b>Name of Provider or Supplier</b>  Dermatology Associates Of Tallahassee Pa	<b>Street Address, City, State</b>  1707 Riggins Rd, Tallahassee, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An onsite announced CLIA recertification survey was conducted at Dermatology Associates of Tallahassee PA, a laboratory in Tallahassee, FL, on 02/02/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D5209</b>	<p><b>PERSONNEL COMPETENCY ASSESSMENT POLICIES</b> CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: 49973 Based on review of the Employee Training and Evaluation Standard Operating Procedure (SOP) and interview with the Laboratory Manager (LM), the laboratory failed to have an approved policy or documentation for the competency assessment of testing personnel (TP) for 2022-2024. The findings include: During a review of employee trainings, there was no documentation to to verify that TP #B received any competency assessments during 2022 or 2023. During an interview on 02/02/24 at 11:15 a.m., the LM verified that TP #B had never been competency assessed. Upon review of the laboratory's Employee Training and Evaluation SOP (last approved by the Laboratory Director in 2024), there was no competency assessment processes for testing personnel listed.</p>