

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0271087	(X3) Date Survey Completed 06/03/2026
Name of Provider or Supplier Pensacola Pediatrics Pa	Street Address, City, State 4951 Grande Dr, Pensacola, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Pensacola Pediatrics PA on 06/03/2026-06/09/2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows:
D3033	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)(i)</p> <p>(a)(3)(i) Records of test system performance specifications that the laboratory establishes or verifies under 493.1253 for the period of time the laboratory uses the test system but no less than 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to establish and implement a written record retention policy. 1.) Review of the "Laboratory's Policy and Procedures," signed by the Laboratory Director (LD) 2/11/26, revealed the absence of a documented record retention policy. 2.) During an interview conducted on 6/3/26 at 12:15 p.m., TP#A confirmed that the laboratory did not have a record retention policy in place. As a result, the laboratory could not demonstrate compliance with CLIA requirements for the retention of laboratory records.</p>