

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0273818	<b>(X3) Date Survey Completed</b>  09/27/2019
<b>Name of Provider or Supplier</b>  Dermatology Group Pa,The	<b>Street Address, City, State</b>  2850 Morningside Dr, Mount Dora, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Recertification survey was conducted on September 10, 2019. The Dermatology Group PA clinical laboratory was found not in compliance with 42 CFR 493, requirements for clinical laboratories.
<b>D5209</b>	<p><b>PERSONNEL COMPETENCY ASSESSMENT POLICIES</b> CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to document competency assessment on 7 out of 7 moderate complexity Testing Personnel from September 2017 to September 2019 (A, B, C, D, E, F &amp; G). Findings: Review of the annual competency records showed that the laboratory failed to have documentation of competency assessments for the moderate complexity Testing Personnel performing the KOH (potassium hydroxide preparation) and Scabies testing. During an interview on 9/27/19 at 11:12 PM, the Practice Administrator acknowledged that they had not performed competencies on the moderate complexity testing personnel performing the KOH and Scabies testing.</p>