

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0276388	(X3) Date Survey Completed 04/17/2019
Name of Provider or Supplier Pediatrics In Brevard Pa	Street Address, City, State 134 S Woods Dr, Rockledge, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to receive a score of at least 80% in the proficiency testing (PT) for bacteriology for the 3rd testing event in 2018. Findings: Review of the laboratory's proficiency testing with American Association of Bioanalysts (AAB) results showed that the laboratory received a score of 70% for bacteriology for the 3rd testing event in 2018. During an interview on 4/17 /19 at 10 AM, Testing Personnel A acknowledged that they had not received at least 80% for bacteriology for the 3rd testing event in 2018.</p>