

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0278214	<b>(X3) Date Survey Completed</b>  03/12/2019
<b>Name of Provider or Supplier</b>  Aqua Dermatology Of Florida, Pa	<b>Street Address, City, State</b>  603 N Flamingo Rd Ste 350, Pembroke Pines, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6046</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory's Technical Consultant failed to perform competency evaluations on testing personnel who performed potassium hydroxide (KOH) and Scabies testing for 2 out of 4 moderate level testing personnel (A and B) in 2018 Findings: Review of the "Laboratory Personnel Report" signed by the laboratory director on 3/12/19 showed that there are four moderate level testing personnel. Review of employee competencies revealed that there was no competency evaluations on moderate Testing Personnel A and B for 2018 During an interview on 1 /12/19 at 9:09 AM, the Clinical Laboratory Administrator acknowledged that a competency evaluation was not performed in 2018 on moderate Testing Personnel A and B.</p>