Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D0278612	**(X3) Date Survey Completed** 07/12/2024
Name of Provider or Supplier Urology Center Of Florida Pa	**Street Address, City, State** 550 Sw 3rd St Ste 305, Pompano Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	A recertification survey conducted from 07/09/2024 to 07/12/2024 found the UROLOGY CENTER OF FLORIDA PA clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D3031	RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3) Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years. This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to retain daily maintenance records for Histology area for one out of two years reviewed. Findings included: Review of Daily Maintenance log for Miles Scientific Vacuum Infiltration Processor (VIP) E300, Tissue Tek dispensing console, Microtome, water bath, room and humidity log, daily maintenance for Hematoxylin and Eosin for 2023 and 2024 revealed that records for 2023 were missing, The annual testing volume for the laboratory based on the Form CMS -116 (03/240 was 80 patients. During an interview on 07/09/2024 at 11:30 AM, the Histology technician confirmed those records were missing and explained that in 2023 there was another person managing the Histology Section and they failed to give them the records for 2023.