Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D0278612	**(X3) Date Survey Completed** 06/08/2026
Name of Provider or Supplier Urology Center Of Florida Pa	**Street Address, City, State** 550 Sw 3rd St Ste 305, Pompano Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An announced CLIA recertification survey was conducted at UROLOGY CENTER OF FLORIDA PA from 05/13/2026 to 06/08/2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiency cited as follows:
D5473	CONTROL PROCEDURES CFR(s): 493.1256(e)(2)(g) (e)(2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to have the documentation of the acceptability of Hematoxylin & Eosin (H&E) staining materials for intended reactivity used for non-gynecologic slides preparations for 174 patients from 1/31/2025 to 05/30/2025. Findings included: 1-Review of the "PATIENT SPECIMEN LOG CYTOLOGY SPECIMENS" log for 2025, revealed that the laboratory tested patients as follows: 1/31/2025 (12 patients), 02/07/2025 (10 patients), 02/14/2025 (7 patients), 02/21/2025 (12 patients), 02/28/2025 (6 patients), 03/07/2025 (13 patients), 03/14/2025 (10 patients), 03/21/2025 (11 patients), 03/28 /2025 (11 patients), 04/04/2025 (12 patients), 04/11/2025 (16 patients), 04/25/2025 (24 patients), 05/02/2025 (13 patients), 05/09/225 (6 patients), 05/16/2025 (7 patients), 05/30/2025 (4 out of 16 patients). 2-Review of "STAINS QUALITY CONTROL" log for 2025, revealed that, the laboratory failed to have documentation of the acceptability of the H&E stain for the non-gynecological slides received from 01 /31/2025 to 05/16/2025 and for four patients of 05/30/2025. During a telephonic interview on 05/18/2026 at 12:56 PM the Testing Personnel confirmed that the laboratory failed to have the acceptability of Hematoxylin & Eosin (H&E) staining materials for intended reactivity used for non-gynecologic slides preparations for 174 patients from 1/31/2025 to 05/30/2025.