

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0284855	(X3) Date Survey Completed 01/25/2018
Name of Provider or Supplier Unilab Of Dade Inc	Street Address, City, State 2145 W Davie Blvd Ste 106, Fort Lauderdale, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on personnel records review and interview with laboratory owner, the laboratory failed to do annual competency evaluations for general supervisor 2 out of 2 from 2016 and 2017. Findings include: Personnel records for 2016 and 2017 showed the laboratory no documented annual competencies for General Supervisor On 01/25 /2018 at 10:30 am, the laboratory owner confirmed that the laboratory did not document annual competencies for the general supervisor for 2016 and 2017.</p>