Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	10D0287446	01/08/2019
Name of Provider or Supplier	**Street Address, City, State**	
Mark S Block Dpm	660 Glades Rd Ste 120, Boca Raton, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on record review and interview with testing personnel (TP) B, the laboratory failed to document the annual competency assessment on 1 out of 2 testing personnel for 2017 and 2018 period. Findings include: Review of employee documentation showed that the laboratory failed to have documentation of annual competency assessment on 2017 and 2018 for TP B. During an interview on 01/08/2019 at 12:30 PM, with TP B, she confirmed that there was no competency assessment documented for the period of reference.