

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0291948	(X3) Date Survey Completed 05/24/2023
Name of Provider or Supplier Advanced Urology Institute	Street Address, City, State 1401 Pasadena Avenue South Suite 4, Saint Petersburg, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification survey was conducted at Advanced Urology Institute on 05/24/23. The laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Office Manager, the laboratory failed to have the correct address of the laboratory location where the urinalysis testing was performed for 5 (Patient #1, #2, #3, #4, #5) out of 5 patient reports reviewed for 2 of 2 years (2021-2023). Findings included: Review of patient reports for Patient #1 (date of service [DOS] 2/16/21), #2 (DOS 1/11/22), #3 (DOS 3/15/22), #4 (DOS 4/28/23), and #5 (DOS 5/23/23) revealed all of the reports had an incorrect address for where the urine specimen testing was performed. On 05/24/23 at 10:25 am, the Office Manager confirmed the address on the 5 patient reports was the laboratory's previous location.</p>