

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0294956	(X3) Date Survey Completed 07/10/2019
Name of Provider or Supplier Bradenton Dermatology Pa	Street Address, City, State 2919 26th St W, Bradenton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Bradenton Dermatology on 07/10/2019. The laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Laboratory Director, the laboratory failed to verify the accuracy of testing twice a year for 2 of 2 years (2017-2018) reviewed for Mycology, Parasitology, and Virology. Findings Included: Review of quality assurance peer review revealed no peer reviews were completed in 2017 for KOH (Potassium hydroxide) testing in Mycology, Scabies testing in Parasitology, and Tzanck testing in Virology. In addition, in 2018 peer review was only conducted one time for Scabies (01/18/18) and Tzanck testing (11/15/18). On 07/10/2019 at 3:30 PM, the Laboratory Director stated he thought the peer reviews for KOH, Scabies, and Tzanck had been performed in 2017. For 2018 the Laboratory Director stated, the laboratory only had 1 positive patient for the Scabies and the Tzanck testing.</p>