

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0663854	(X3) Date Survey Completed 05/12/2026
Name of Provider or Supplier Pediatric Associates Inc	Street Address, City, State 9611 W Broward Blvd, Plantation, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at PEDIATRIC ASSOCIATES INC from 05/08/2026 to 05/12/2026 . The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiency cited as follows:
D2007	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to examine or test Proficiency Testing (PT) samples with the laboratory's regular patient workload (testing Personnel (TP)) that routinely perform the testing in the laboratory for four out of four events reviewed for Chemistry specialty. Findings included: 1-Review of FORM CMS-209 signed and dated by the Laboratory Director (LD) on 05/08/2026 revealed the laboratory had six TP listed (TP#1, TP#2, TP#3, TP#4, TP#5 and TP#6). 2-Review of the procedure manual signed by the Laboratory Director (LD) in policy titled, "LAB 410 Proficiency Testing" revealed that stated under "SCOPE" "All laboratory staff are to review and participate in the proficiency program. Any employee who is working in the lab and performing moderate complex patient testing must participate in Proficiency Testing" 3-Review of American Proficiency Institute (API) PT attestation statement records for Chemistry specialty for the following events 2025 (1st, 2nd and 3rd) and 2026 (1st event) revealed that TP#2 failed to sign an attestation in 2025 three events and first event of 2026. 4-During an interview on 05/08/2026 at 12:20 PM, the Lead Medical Assistant confirmed that TP#2 failed to participate in PT in Chemistry specialty for the events listed above.</p>