

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0665433	(X3) Date Survey Completed 01/08/2020
Name of Provider or Supplier Pardo Forstot Baca & Alboukrek Pa	Street Address, City, State 1050 Nw 15 St Ste 212a, Boca Raton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced recertification survey was conducted on 1/8/20 at Pardo Forstot Baca & Alboukrek, PA, a clinical laboratory in Boca Raton, Florida. The laboratory was not in compliance with CFR Part 493, requirements for clinical laboratories.
D2099	<p>ENDOCRINOLOGY CFR(s): 493.843(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory failed to attain 80 percent for free thyroxine on one testing event. Findings include: Review of proficiency testing records for the past two years revealed that on the third testing event of 2018, the laboratory received a score of 40% for free thyroxine. During an interview with the general supervisor at 10:30 a.m., she confirmed that they had gotten an unsatisfactory score.</p>
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory failed to attain 80 percent for hematocrit on one testing event. Findings include: Review of proficiency testing records for the past two years revealed that on the third testing event of 2019, the laboratory received a score of 40% for hematocrit. During an</p>

interview with the general supervisor at 10:30 a.m., she confirmed that they had gotten an unsatisfactory score.