

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0669049	<b>(X3) Date Survey Completed</b>  05/17/2021
<b>Name of Provider or Supplier</b>  University Of Miami Hepatology Diagnostic	<b>Street Address, City, State</b>  1600 Nw 10th Ave Rm 7163, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey was completed on 5/17/2021 at UNIVERSITY OF MIAMI HEPATOLOGY DIAGNOSTIC LABORATORY. The clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D2076</b>	<p>GENERAL IMMUNOLOGY CFR(s): 493.837(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to receive a passing score of at least 80% in General Immunology, for the 3rd event of 2020 with the College of American Pathologist Proficiency (CAP). Findings Included : Review of CAP 2020 Proficiency Testing revealed 0% for Hepatitis B surface (HBsAg), 0% for Hepatitis B core antibody (Anti-Hbc) and 33% for General Immunology overall during the 3rd proficiency testing event in 2020. During an interview on 5/17/2021 at 3:51 pm, the office supervisor confirmed that the laboratory received a score of 33% for general Immunology, for the 3rd event in 2020.</p>