

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0674212	<b>(X3) Date Survey Completed</b>  03/21/2018
<b>Name of Provider or Supplier</b>  Saul Lipsman Dpm Md Pa	<b>Street Address, City, State</b>  6894 Lake Worth Rd Ste 102, Lake Worth, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5477</b>	<p>CONTROL PROCEDURES CFR(s): 493.1256(e)(4)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (4) Before, or concurrent with the initial use-- (e)(4)(i) Check each batch of media for sterility if sterility is required for testing; (e)(4)(ii) Check each batch of media for its ability to support growth and, as appropriate, select or inhibit specific organisms or produce a biochemical response; and (e)(4)(iii) Document the physical characteristics of the media when compromised and report any deterioration in the media to the manufacturer. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory did not check each new batch or lot of media to produce the appropriate biochemical response by using a positive and negative control. Findings include: Review of fungal culture quality control records and interview with the office manager at 10:10 a.m. on 03/21/18 revealed that the laboratory tested each lot of the media they used for fungal cultures with a positive control, but did not use a negative control to ensure the biochemical response, color change, only with appropriate organisms.</p>