

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0674401	(X3) Date Survey Completed 07/28/2022
Name of Provider or Supplier Medical Center Clinic Dermatology	Street Address, City, State 8333 N Davis Hwy, Pensacola, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on July 28, 2022. Medical Center Clinic Dermatology clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5291	<p>GENERAL LABORATORY SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1239(a)</p> <p>The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and, when indicated, correct problems identified in the general laboratory systems requirements specified at 493.1231 through 493.1236.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Laboratory Director (LD) and Testing Personnel (TP) #A, the laboratory failed to perform and document Quality Assessment (QA) activities for the the years 2020, 2021, and from January to July of 2022. Findings include: Based on record review of the QA policy and procedure, it was revealed no QA activity had been performed and documented for the years 2020, 2021, and from January to July 2022. It was confirmed during interview on July 28, 2022 at 11:35 a.m. with the LD and TP #A, no QA activities had been performed or documented.</p>