

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0677474	<b>(X3) Date Survey Completed</b>  06/12/2019
<b>Name of Provider or Supplier</b>  Minars Dermatology Llc	<b>Street Address, City, State</b>  4060 Sheridan St Ste C, Hollywood, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Minars Dermatology Llc on 6/12/2019. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D5217</b>	<p><b>EVALUATION OF PROFICIENCY TESTING PERFORMANCE</b> CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with Testing Personnel (TP) # 2, the laboratory failed to ensure the twice a year accuracy verification testing for Mycology for 2 (#6 and #3) in 2018 and 1 (#6) for 2017 out of 5 TP of the KOH test, subspecialty of Mycology; for 2 out of 2 (2017 and 2018) years reviewed. Findings include: Review of Proficiency Testing records, revealed: - No documentation of the peer review for KOH Test, Mycology for TP # 6 for 2017 and 2018. - No documentation of 2017 peer review for KOH test, Mycology for TP # 3 for 2018 During an interview on 06/12 /2019 at 2:30 PM, the TP # 2 confirmed that the laboratory had no documentation of the accuracy verification for the TP of reference during the years listed above.</p>