

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0678337	<b>(X3) Date Survey Completed</b>  10/05/2020
<b>Name of Provider or Supplier</b>  Center For Prevention	<b>Street Address, City, State</b>  921 North Davis Street Bldg A Suite 251 Mc-68, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	At the time of the announced, on-site recertification survey, Duval County Public Health Department was found to be in non-compliance with the CLIA laboratory regulatory requirements of 42 CFR 493.
<b>D5217</b>	<p><b>EVALUATION OF PROFICIENCY TESTING PERFORMANCE</b> CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the facility failed to ensure that at least twice annually, the laboratory verified the accuracy of tests performed in the subspecialties of Mycology and Parasitology for two of two years reviewed. The findings include: Based on record review on 10/5/20, the laboratory failed to provide documentation showing twice annual verification of accuracy for wet preps and KOH (potassium hydroxide). Interview with Testing Person A on 10/5/20 at 11:30am indicated that verification of accuracy was only performed annually.</p>