

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0681074	(X3) Date Survey Completed 10/18/2018
Name of Provider or Supplier Adam S Plotkin Md Pa D/B/A	Street Address, City, State 5210 Linton Blvd Ste 307, Delray Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the laboratory's proficiency testing records for 2017 and 2018, the laboratory had subsequent unsuccessful performance in proficiency testing for the specialty of mycology. Refer to D2046. Findings include: Review of the American Proficiency Institute (API) proficiency testing records and the review of the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports, on October 18, 2018 at 10:00 AM, showed that the laboratory had unsatisfactory testing scores in the specialty of mycology for three testing events out of five testing events in 2017 and 2018.</p>

<p>D2046</p>	<p>MYCOLOGY CFR(s): 493.827(e)</p> <p>Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on the review of the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports and the laboratory's proficiency testing records, the laboratory had subsequent unsuccessful participation in proficiency testing for the specialty of mycology. Findings include: On October 18, 2108 on or about 10:00 AM the American Proficiency Institute (API) proficiency testing records and the CMS 153 and 155 reports were reviewed. The review showed that the laboratory had unsuccessful proficiency testing scores for three out of five testing events in 2017 and 2018. Due to the laboratory's third failure, the laboratory had a subsequent unsuccessful participation in proficiency testing. Event #1, 2017 Mycology-60% Event #3, 2017 Mycology-60% Event #2, 2018 Mycology-60%</p>
<p>D6076</p>	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on the review of the laboratory's proficiency testing records, the laboratory director failed to ensure that the laboratory maintained a satisfactory score for proficiency testing in the specialty of mycology. Findings include: On October 18, 2018, on or about 10:00 AM, the American Proficiency Institute (API) proficiency records and the Centers for Medicare & Medicaid Service (CMS) 153 and 155 reports were reviewed. The review showed that the laboratory had subsequent unsatisfactory proficiency testing score for the specialty, mycology. The laboratory director is responsible for ensuring that the laboratory maintains successful participation in proficiency testing. Refer to D2046.</p>
<p>D6089</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i)</p> <p>The laboratory director must ensure the proficiency testing samples are tested as required under subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on the review of the laboratory's proficiency testing scores, the laboratory director failed to ensure that the laboratory performed proficiency testing in such a manner as to achieve and maintain successful participation in proficiency testing in the specialty of mycology. Findings Include: The review of the American Proficiency Institute (API) proficiency testing records and the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports on October 18, 2018 on or about 10:00 AM</p>

showed that the laboratory received unsatisfactory proficiency testing scores for three out of three out of five testing events in 2017 and 2018. Due to the laboratory's third failure, the laboratory had a subsequent unsuccessful participation in proficiency testing. The test scores are shown below. Event #1, 2017 Mycology-60 % Event #3, 2017 Mycology-60 % Event #2, 2018 Mycology-60 %