

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0683133	<b>(X3) Date Survey Completed</b>  11/28/2018
<b>Name of Provider or Supplier</b>  Palm Beach County Health Department-Laboratory	<b>Street Address, City, State</b>  1150 45th St Rm L-200, West Palm Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6127</b>	<p> <b>TECHNICAL SUPERVISOR RESPONSIBILITIES</b>                      CFR(s): 493.1451(b)(9)                 </p> <p>                     The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.                 </p> <p>                     This STANDARD is not met as evidenced by:                      Based on record review and interview with laboratory personnel, the technical supervisor did not document a six month competency evaluation for one of two testing personnel hired since the last survey, Findings include: Review of personnel records on 11/28/18 revealed that one of the two technologists hired over the past year did not have any documentation to indicate that a competency evaluation had been done since she was hired 02/05/18. During an interview with her supervisor who was also one of the technical supervisors, he did not recall documenting an evaluation of her.                 </p>