Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	10D0686545	01/25/2018
Name of Provider or Supplier	**Street Address, City, State**	
Hobart K Richey Md Pa	728 The Rialto, Venice, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5200	GENERAL LABORATORY SYSTEMS CFR(s): 493.1230 Each laboratory that performs nonwaived testing must meet the applicable general laboratory systems requirements in 493.1231 through 493.1236, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the general laboratory systems and correct identified problems specified in 493.1239 for each specialty and subspecialty of testing performed. This CONDITION is not met as evidenced by: Based on record review and interview with the Office Manager the laboratory failed to perform peer review every 6 months in Virology for 1 (2017) out of 2 (2016-2018) years reviewed and in Parasitology for 2 out of 2 (2016-2018) years reviewed (See D5217).
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on record review and interview with the Office Manager the laboratory failed to perform peer review every 6 months in Virology for 1 (2017) out of 2 (2016-2018) years reviewed and in Parasitology for 2 out of 2 (2016-2018) years reviewed. This is a repeat deficiency from the 01/27/14 recertification survey. Findings Included: Review of peer review for Virology found Tzank testing peer review performed 01/06 /16, 05/20/16, 06/07/16, 08/25/16, 10/05/16, 01/13/17, 01/15/17, 01/23/17, 03/30/17,

	(10 month gap between these dates) 01/18/18, 01/19/18. Review of peer review for Parasitology found no peer reviews for the Scabies testing. During an interview on 01 /25/18 at 2:00 PM the Office Manager confirmed that there were no other peer reviews. This is a repeat deficiency from the 01/27/14 recertification survey. Review of the plan of correction (signed by the Laboratory Director 02/10/14) revealed that "The laboratory will perform a minimum of one QA peer review for Tzank, scabies and KOH every six months." It also stated "The laboratory director has been made aware of this deficiency and will be responsible for seeing that it does not recur."