

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0686545	<b>(X3) Date Survey Completed</b>  04/08/2024
<b>Name of Provider or Supplier</b>  Hobart K Richey Md Pa	<b>Street Address, City, State</b>  728 The Rialto, Venice, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An on-site announced CLIA recertification survey was conducted at Hobart K Richey MD PA on 04/03/2024 - 04/08/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
<b>D5417</b>	<p><b>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT</b> CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview with the office manager, the laboratory failed to ensure the potassium hydroxide that was used for mycology (fungi) and parasitology (scabies) testing was not expired prior to performing 27 patient tests from 11/03/23- 04/02/24. Findings included: A tour of the laboratory on 04/03/2024 at 03:00 PM, revealed the expiration date was 10/19/2023 on the bottle of potassium hydroxide (Lot number 1292) used for mycology (fungi) and parasitology (scabies) testing. A review of the patient fungi and scabies result log revealed 27 tests had been performed after the 10/19/2023 expiration date. The testing was performed on 11/03/23 (1 test), 11/14/23 (1 test), 11/15/23 (1 test), 11/21/23 (3 tests), 12/4/23 (1 test), 12/5/23(1 test), 12/7/23 (3 tests), 01/02/24 (2 tests), 01/03/24 (1 test), 01/10/24 (1 test), 01/29/24 (3 test), 01/31/24 (1 test), 02/06/24 (1 test), 02/21/24 (1 test), 03/13/24 (1 test), 03/15/24 (1 test), 03/28/24 (1 test), 04/01/2024 (1 test), and 04/02/24 (2 test). On 04/03/2024 at 3:05 PM, the Office Manager confirmed the potassium hydroxide was expired and patients were tested.</p>