

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0686969	(X3) Date Survey Completed 05/20/2022
Name of Provider or Supplier South Florida Dermatology Group, Inc	Street Address, City, State 401 Coral Way Ste 207, Coral Gables, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 05/20/2022 found the VITOR F WEINMAN MD clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to ensure twice a year accuracy verification of potassium hydroxide (KOH) preps for one out of two years reviewed (2020-2021). Findings include: -Review of Quality Control policy, in the Interpretation Quality Control Section revealed that the laboratory will do an interpretation verification twice a year for KOH test. -Review of "KOH LOG" revealed that the laboratory performed one verification on 03/12/2020, no records found of another verification in 2020. During an interview on 05/20/22 at 11:00 AM, the office manager confirmed that the laboratory failed to perform twice a year accuracy verification for the KOH preps during 2020.</p>