

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0688834	(X3) Date Survey Completed 08/24/2026
Name of Provider or Supplier Alvaro J Dangond Md Pa	Street Address, City, State 7001 Sw 97th Ave Ste 201, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at ALVARO J DANGOND MD PA on August 26, 2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiency cited as follows:
D5437	CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(a) (a)Unless otherwise specified in this subpart, for each applicable test system the laboratory must perform and document calibration procedures-- (a)(1) Following the manufacturer's test system instructions, using calibration materials provided or specified, and with at least the frequency recommended by the manufacturer; (a)(2) Using the criteria verified or established by the laboratory as specified in 493.1253(b) (3)-- (a)(2)(i) Using calibration materials appropriate for the test system and, if possible, traceable to a reference method or reference material of known value; and (a) (2)(ii) Including the number, type, and concentration of calibration materials, as well as acceptable limits for and the frequency of calibration; and (a)(3) Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification. This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to perform calibration verification procedures of the Medonic M Series Analyzer at least every 6 months in 2026. Findings included: 1- Review of the Medonic M series Hematology Manual, revealed that in the Calibration section stated: "Calibration must be performed upon set up of the instrument and then at a minimum of every 6 months." - Review of the Medonic M Series Analyzer calibration records revealed that the laboratory performed calibrations during 2025 and the last record was 11/06/2025. The laboratory had no records of calibration in 2026. During an interview on 08/24 /2061 at 12:00 PM, the Laboratory Director confirmed that the laboratory failed to perform instrument calibration every 6 months in 2026.