

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0689203	(X3) Date Survey Completed 01/28/2020
Name of Provider or Supplier Ruth C Schobel Md Pa	Street Address, City, State 7480 Fairway Dr Ste 202, Miami Lakes, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on January 28, 2020. Ruth C Schobel MD PA clinical laboratory was found not in compliance with 42 CFR 493, requirements for clinical laboratories.
D2123	<p>HEMATOLOGY CFR(s): 493.851(c)</p> <p>Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to participate in proficiency testing that resulted in a score of zero (0) percent in Hematology /Coagulation for 1(2018 3rd) out of 6 (2018 1st, 2nd and 3rd, 2019 1st, 2nd and 3rd) events. Findings: Review of the American Proficiency Institute (API) records for the 2018 Hematology/Coagulation 3rd event showed the laboratory received a score of 0% for all analytes to be tested. Notes on the API "Performance Summary" sheet stated "Failure to Participate". During an interview on 1/28/19 at 3:30 PM, the Testing Personnel C acknowledged that the laboratory failed to transmit the proficiency testing results to API for the 3rd event in 2018 before the deadline.</p>