

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0694170	<b>(X3) Date Survey Completed</b>  03/08/2018
<b>Name of Provider or Supplier</b>  Rene L Lopez-Guerrero Md Pa	<b>Street Address, City, State</b>  3445 Nw 7th St, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2020</b>	<p><b>BACTERIOLOGY</b> CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of College of American Pathologist (CAP) Proficiency Testing (PT) results and interview with the laboratory director, the laboratory failed to have satisfactory results in bacteriology in 1 out of 1 event. The findings include: Review of CAP PT records for year 2016 showed a score of 67 % for Bacteriology in 2016 2nd event. During an interview with the laboratory director at 12:00 pm on March 8th 2018 he confirmed that the laboratory failed PT for bacteriology specialty in 2016 second event.</p>
<b>D2122</b>	<p><b>HEMATOLOGY</b> CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of College of American Pathologist (CAP) Proficiency Testing (PT) results and interview with the laboratory director, the laboratory failed to have satisfactory results in hematology in 1 out of 1 event. The findings include: Review of CAP PT records for year 2017 showed a score of 76 % for Hematology in 2017 1st event. During an interview with the laboratory director at 12:00 pm on March 8th 2018 he confirmed that the laboratory failed PT for hematology specialty in 2017 first event.</p>