

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0697355	(X3) Date Survey Completed 12/04/2019
Name of Provider or Supplier G Alexander Carden Md Pa & David W Dodson Md Pa	Street Address, City, State 1411 N Flagler Dr Ste 7900, West Palm Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced initial certification survey was conducted on 12/4/19 at G. Alexander Carden MD PA and David Dodson MD PA, a clinical laboratory in West Palm Beach, Florida. G. Alexander Carden MD PA and David Dodson MD PA is not in compliance with Code of Federal Regulations (CFR), Part 493, requirements for clinical laboratories.
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory failed to attain a satisfactory score for one analyte on the second testing event of 2019. The findings included: Review of that past two years of proficiency testing records on 12/4/19 revealed that the laboratory received a score of 60% for albumin on the second testing event of 2019. During an interview at 12:10 p.m. on 12/4/19, the laboratory consultant confirmed that they had received an unsatisfactory score, and that they had taken corrective action.</p>