

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 10D0707800 | (X3) Date Survey Completed 02/15/2019 |
| Name of Provider or Supplier American Health S Llc DbA American Health | Street Address, City, State 15712 Sw 41 St Suite 16, Davie, FL | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D2020 | <p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologist (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory failed to get at least 80 % for 1 (2nd event of 2018) out of 6 events for Bacteriology. Findings include: Review of CAP proficiency records for 2017 and 2018 revealed that the laboratory had unacceptable results for 2 out of 5 bacterial identification samples receiving a score of 60 % and resulting in an overall score of 73 % in Bacteriology 2nd event of 2018. During an interview on 02/13/2019 at 3:30 PM, the TS confirmed that the laboratory failed the event of reference.</p> |
| D2087 | <p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologist (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory failed to score at least 80 % on Creatine Kinase (CK) for 1 (3rd event of 2017) out of 6 events for Routine Chemistry reviewed. Findings include: Review of CAP proficiency records revealed a score of 40% for CK in the 3rd event of 2017. During an interview on 02/13 /2019 at 3:30 PM, the TS confirmed the failed proficiency testing score.</p> |

| | |
|---------------------|---|
| <p>D2109</p> | <p>TOXICOLOGY CFR(s): 493.845(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologist (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory failed to score at least 80 % for analyte Theophylline in 1 (1st event of 2017) out of 6 events for Toxicology reviewed. Findings include: Review of CAP proficiency records revealed a score of 60% for Theophylline in the 1st event of 2017. During an interview on 02 /13/2019 at 3:30 PM, the TS confirmed the proficiency testing score.</p> |
| <p>D2121</p> | <p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologist (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory failed to score at least 80 % on Blood Cell Identification (BCI) for 1 (3rd event of 2018) out of 6 events for Hematology reviewed. Findings include: Review of CAP proficiency records revealed a score of 60% for BCI in the 3rd event of 2018. During an interview on 02 /13/2019 at 3:30 PM, the TS confirmed the failed proficiency testing score.</p> |
| <p>D2122</p> | <p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College American Pathologist (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory received an unsatisfactory score in 1 (3rd event of 2017) out of 6 events for Hematology reviewed. Findings include: Review of CAP proficiency records for 2017 and 2018 revealed that the laboratory received a 0 % score for Hematocrit and Hemoglobin resulting in an overall score of 75 %. During an interview on 02/13/2019 at 3:30 PM, the TS confirmed that the laboratory failed the event of reference.</p> |