

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0710944	(X3) Date Survey Completed 01/31/2023
Name of Provider or Supplier Kevin H Lapoff Dpm	Street Address, City, State 6422 Lake Worth Rd, Lake Worth, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced recertification survey was conducted from 01/24/2023 to 01/31/2023 at KEVIN H LAPOFF DPM, a clinical laboratory in Lake Worth, Florida. KEVIN H LAPOFF DPM clinical laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D3033	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)(i)</p> <p>In addition, the laboratory must retain records of test system performance specifications that the laboratory establishes or verifies under 493.1253 for the period of time the laboratory uses the test system but no less than 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory director, the laboratory failed to retain daily room temperature records for 18 out of 24 months reviewed (2021 and 2022). Findings include: Review of laboratory temperature logs revealed that the laboratory failed to have room temperature logs during 2021 and during 2022 from January to June. During an interview on 01/24/2023 at 10:40 a.m., the laboratory director confirmed that the laboratory failed to have onsite the temperature records of reference.</p>