

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0714948	(X3) Date Survey Completed 06/22/2018
Name of Provider or Supplier Brian D Ardel Md Pa	Street Address, City, State 3417-D Tamiami Tr, Port Charlotte, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the laboratory's proficiency testing records for 2017 and 2018, the laboratory had subsequent unsuccessful performance in proficiency testing in the specialty of hematology. Refer to D2130. Findings include: Review of the American Proficiency Institute (API) proficiency testing records and the review of the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports, on June 25, 2018 at 11:00 AM, showed that the laboratory had unsatisfactory testing scores for the analyte, white blood cell (WBC) differential for three out of five testing events in 2017 and 2018.</p>

<p>D2130</p>	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on the review of the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports and the laboratory's proficiency testing records, the laboratory had subsequent unsuccessful participation in proficiency testing in the specialty of hematology. Findings include: On June 25, 2018, on or about 11:00 AM the American Proficiency Institute (API) proficiency testing records and the CMS 153 and 155 reports were reviewed. The review showed that the laboratory had unsuccessful proficiency testing scores for the analyte, white blood cell (WBC) differential, for three out of five testing events in 2016, 2017, and 2018 as shown below. Event #3, 2016 WBC differential-60% Event #2, 2017 WBC differential-0% Event #1, 2018 WBC differential-73%</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on the review of the laboratory's proficiency testing records, the laboratory director failed to ensure that the laboratory maintained a satisfactory score for proficiency testing in the specialty of hematology. Findings include: On June 25, 2018 on or about 11:00 AM, the American Proficiency Institute (API) proficiency records and the Centers for Medicare & Medicaid Service (CMS) 153 and 155 reports were reviewed. The review showed that the laboratory had subsequent unsatisfactory proficiency testing score for the analyte, white blood cell (WBC) differential, in the specialty of hematology. The laboratory director is responsible for ensuring that the laboratory maintains successful participation in proficiency testing. Refer to D2130.</p>
<p>D6016</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on the review of the laboratory's proficiency testing scores, the laboratory director failed to ensure that the laboratory performed proficiency testing in such a manner as to achieve and maintain successful participation in proficiency testing for</p>

the analyte white blood cell (WBC) differential in the specialty of hematology. Findings Include: The review of the American Association of Bioanalysts (API) proficiency testing records and the Centers for Medicare & Medicaid Services (CMS) 153 and 155 reports on April 20, 2017 on or about 10:00 AM showed that the laboratory received unsatisfactory proficiency testing scores for three out of five testing events in 2016, 2017, and 2018 as shown below. Event #1, 2016 WBC differential-60% Event #2, 2017 WBC differential-0% Event #1, 2018 WBC differential-73%