

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0717479	<b>(X3) Date Survey Completed</b>  06/22/2020
<b>Name of Provider or Supplier</b>  Gulf Coast Pathologists Llc	<b>Street Address, City, State</b>  2631 McCormick Dr Suite 101, Clearwater, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Gulf Coast Pathologists LLC on 06/22/20. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Cytotechnologist the laboratory failed to verify the accuracy of Cytology testing at least twice a year for 1 year (2019) out of 2 years reviewed (2018-2020). Findings Included: Review of Cytology peer reviews revealed that it was completed 06/18/20, 11/20/18, and 09/19/18. There were no reviews done in 2019. Interview on 06/22/20 at 11:30 AM with the Cytotechnologist confirmed that the reviews were not performed in 2019.</p>