

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0719864	(X3) Date Survey Completed 01/30/2018
Name of Provider or Supplier Womens Ob/Gyn Center Of Countryside Inc	Street Address, City, State 28960 Us Hwy 19 N Ste 110, Clearwater, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2153	<p>ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(a)</p> <p>Failure to attain a score of at least 100 percent of acceptable responses for each analyte or test in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of AAB (American Association of Bioanalysts) proficiency testing and interview with the Office Manager, the laboratory failed to get a 100% on proficiency testing for 1 (3rd testing event 2016) out of 6 testing events (1st, 2nd, 3rd testing event 2016 and 2017) reviewed. Findings Included: Review of AAB proficiency testing revealed that in the 3rd testing event of 2016 the laboratory scored 60% for Rh typing. During an interview on 01/30/18 at 2:00 PM the Office Manager confirmed the proficiency testing failure.</p>