

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0719864	<b>(X3) Date Survey Completed</b>  04/29/2024
<b>Name of Provider or Supplier</b>  Womens Ob/Gyn Center Of Countryside Inc	<b>Street Address, City, State</b>  28960 Us Hwy 19 N Ste 110, Clearwater, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Womens OB/GYN Center of Countryside Inc on 4/29/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
<b>D5203</b>	<p><b>SPECIMEN IDENTIFICATION AND INTEGRITY</b> CFR(s): 493.1232</p> <p>The laboratory must establish and follow written policies and procedures that ensure positive identification and optimum integrity of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to have a written policy that ensured positive identification of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results for 2 of 2 years (2022 - 2024). Findings included: 1. Review of the Specimen Collection and Handling policy revealed specimens were to be labeled. The policy did not include what information was needed to ensure positive identification and the integrity of patients' specimens. 2. Review of patient logs for 12/2022, 7/2023, and 2 /2024 revealed only patient names were used for labeling. 3. Interview with Testing Person #A confirmed on 4/29/2024 at 10:40 AM only patients' names were used for specimen identification, and the policy did not include what information to include to ensure positive identification.</p>