

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 10D0722984	<b>(X3) Date Survey Completed</b> 04/03/2026
<b>Name of Provider or Supplier</b> Advanced Dermatology And Cosmetic Surgery	<b>Street Address, City, State</b> 1639 Atlantic Blvd Ste 100, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Advanced Dermatology and Cosmetic Surgery on 4/3/2026. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D3001</b>	<p><b>FACILITIES</b> CFR(s): 493.1101(a)(1)</p> <p>(a) The laboratory must be constructed, arranged, and maintained to ensure the following: (a)(1) The space, ventilation, and utilities necessary for conducting all phases of the testing process.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview, the laboratory failed to ensure that the ventilation system was maintained to properly remove chemical vapors for the testing performed, when the facility lacked a chemical fume hood required by the manufacturers of the histopathology reagents (100% Dehydrant and Eosin-Y with Phloxine). Findings include: 1. During an observation of the laboratory on 4/3/2026, it was noted that the facility lacked a chemical fume hood. An EpreDia Linistat Linear Stainer was observed situated on a workbench with multiple open reagent containers used for staining patient specimens. There was no local exhaust ventilation or fume hood present above or around this equipment. 2. The following chemical reagents, used for histopathology testing, were stored in the facility's flammable cabinet on 4/3/2026: 100% Dehydrant (Manufacturer: Cardinal Health) Eosin-Y w/ Phloxine (Manufacturer: EpreDia) 3. A review of the Safety Data Sheets (SDS) for both the 100% Dehydrant and the Eosin-Y w/ Phloxine on 4/3/2026 revealed manufacturer instructions regarding ventilation requirements: The SDS for 100% Dehydrant stated under Section 7 (Handling) and Section 8 (Engineering Measures): "Use only under a chemical fume hood." The SDS for Eosin-Y w/ Phloxine stated under Section 7 (Handling) and Section 8 (Engineering Measures): "Use only under a chemical fume</p>

hood." 4. In an interview on 4/3/2026 at 9:30 AM, the Mohs Technician confirmed the laboratory did not have a chemical fume hood. The Mohs Technician further stated they were aware the safety data sheets for the dehydrant and Eosin stains required the use of a chemical fume hood to protect personnel from chemical vapors during the testing process.