

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0722985	(X3) Date Survey Completed 07/13/2023
Name of Provider or Supplier Advanced Dermatology And Cosmetic Surgery	Street Address, City, State 1893 Kingsley Avenue, Orange Park, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	At the time of the announced onsite recertification survey, Advanced Dermatology and Cosmetic Surgery, was found to not be in compliance with the CLIA laboratory requirements of 42 CFR 493.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on review of the reagent log and interview, the laboratory failed to ensure the Eosin Y stain used for Mohs testing was not expired prior to patient testing for 11 days in 2022. The findings include: The review of the laboratory form titled "Daily Quality Control Slide" showed patient testing was performed on 4/6/22, 4/13/22, 4/20/22, 4/27/22, and 5/4/22 using Eosin Y stain with lot number 96733 and an expiration date of 3/31/22. The review of the laboratory Mohs log showed approximately 78 patients were tested over the 5 days the expired Eosin Y was used. Further review of the form titled "Daily Quality Control Slide" showed patient testing was performed on 10/5/22, 10/12/22, 10/19/22, 10/26/22, 11/2/22, and 11/9/22 using Eosin Y stain with lot number 1050777 and expiration date of 9/2022. The review of the laboratory Mohs log showed approximately 107 patients were tested over the 6 days the expired Eosin Y was used. During the interview with the Area Clinical Manager on 7/13/23 at 9:05am, it was confirmed the documentation showed expired Eosin Y stain was used.</p>