

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0865961	(X3) Date Survey Completed 02/24/2022
Name of Provider or Supplier Dan H Meirson Md Pa	Street Address, City, State 1166 W Newport Center Dr Ste 100, Deerfield Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on February 24, 2022. Dan H Meirson MD PA clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to verify accuracy of the reading and interpretation of the Hematoxylin and Eosin (H&E) stain at least twice annually in 2021. Findings: Review of "Quality Assurance Proficiency Testing" showed there were four case reviewed for proficiency testing in 2021. The "Date received by the Mohs technologist And Material received, reviewed by" the Physician on the proficiency testing form was dated 10/20/2021 for all four cases. On 02/24/2022 at 1:40 PM, the Histology Technician acknowledged the four cases were all dated 10/20/2021.</p>