

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0865961	<b>(X3) Date Survey Completed</b>  02/27/2026
<b>Name of Provider or Supplier</b>  Dan H Meirson Md Pa	<b>Street Address, City, State</b>  1166 W Newport Center Dr Ste 100, Deerfield Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at DAN H MEIRSON MD PA on February 27, 2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiency cited as follows:
<b>D5781</b>	<p><b>CORRECTIVE ACTIONS</b> CFR(s): 493.1282(b)(1)</p> <p>(b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(1) Test systems do not meet the laboratory's verified or established performance specifications, as determined in 493.1253(b), which include but are not limited to-- (b)(1)(i) Equipment or methodologies that perform outside of established operating parameters or performance specifications; (b)(1)(ii) Patient test values that are outside of the laboratory's reportable range of test results for the test system; and (b)(1)(iii) When the laboratory determines that the reference intervals (normal values) for a test procedure are inappropriate for the laboratory's patient population.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to document corrective actions (CA) for cryostat temperatures that were not within acceptable temperatures range from 10/01/2025 to 02/25/2026 in 8 out of 19 testing dates. Findings included: 1-Review of policy "EQUIPMENT QUALITY CONTROL-CRYOSTAT", revealed that the laboratory established since July 23, 2004, an acceptable temperature range of -20C To -30C. 2- Review of the Cryostat Log for the period of 10/01/2025 to 02/25/2026, revealed the following dates the temperature was outside of the acceptable range: 10/08/2025, 10/15/2025, 10/29/2025, 11/26/2025, 11/26/2025, 12/03/2025, 01/21/2026, 02/22/2026, with a recorded temperature of -19C.</p>

During an interview on 02/27/2026 at 12:45 AM the risk management consultant confirmed that there was no documentation of the corrective actions for the testing dates that the recorded Cryostat temperature was outside of the acceptable range.