

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0880781	(X3) Date Survey Completed 03/13/2018
Name of Provider or Supplier Donald B Morris Md Pa	Street Address, City, State 1111 7th Ave N Ste 101, Saint Petersburg, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on American Proficiency Institute (API) record review and interview with the Laboratory Director, the laboratory failed to satisfactorily participate in one (2nd event 2017) event out of six (1st, 2nd, 3rd events in 2016 and 2017) events. Findings included: Record review of the American Proficiency Institute (API) Performance Summary included an unsatisfactory score of 50% for the Microbiology 2017 2nd Event for the sub-specialty of Bacteriology. During an interview on 03/13/18 at 12:15 PM, the Laboratory Director confirmed the unsatisfactory score for the Microbiology 2017 2nd Event for the sub - specialty of Bacteriology.</p>