

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0888169	(X3) Date Survey Completed 05/16/2019
Name of Provider or Supplier Planned Parenthood Of Southwest & Central	Street Address, City, State 1425 Creech Rd, Naples, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2010	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(2)</p> <p>The laboratory must test samples the same number of times that it routinely tests patient samples.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory did not test proficiency testing samples the same number of times as it routinely tests patient samples. Findings include: The laboratory routinely tests patients for Rhesus factor (Rh factor) once, but review of proficiency testing records for the past two years on 05/16/2019 revealed that American Proficiency Institute (API) proficiency testing samples for Rh factor were tested more than once. For the first and second testing events of 2018, five testing personnel tested each of the five samples. For the third testing event of 2018, two testing personnel tested samples #11 and 14. For the first testing event of 2019, four testing personnel tested each of the five samples. During an interview with the manager at 11:00 a.m. on 05/16/2019, she confirmed that more than one testing person tested the proficiency test samples.</p>