

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0895195	(X3) Date Survey Completed 05/18/2020
Name of Provider or Supplier Dermatology Associates	Street Address, City, State 350 Nw 76th Dr Ste A, Gainesville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	At the time of the announced, onsite recertification survey, Dermatology Associates was found to NOT be in compliance with the CLIA laboratory requirements of 42 CFR 493. .
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory personnel, the laboratory used a reagent that had exceeded its expiration date. The findings include: During a tour of the laboratory on 5/18/20, one opened container of Toluidine Blue, lot #K17AL1 expiration date 10/31/19, was observed next to the microscope. There were no other containers of Toluidine Blue in the laboratory. During an interview with the laboratory testing person at 1:20pm on 5/18/20, she confirmed that the Toluidine Blue had expired.</p>