

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0895195	(X3) Date Survey Completed 06/19/2024
Name of Provider or Supplier Dermatology Associates	Street Address, City, State 350 Nw 76th Dr Ste A, Gainesville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	At the time of the announced, onsite recertification survey, Dermatology Associates was found to NOT be in compliance with the CLIA laboratory requirements of 42 CFR 493.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to send histopathology patient slides for peer review more than once annually in 2022. Findings include: The record review of documented peer review for 2022 showed that patient slides were sent on 6/9/2022 to a Dermatopathologist to verify accuracy of results. There was no other documented peer review for 2022. The interview with Testing Person E on 6/19/24 at 10:00am confirmed that slides were only sent out once for peer review in 2022.</p>