

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0899197	<b>(X3) Date Survey Completed</b>  11/26/2018
<b>Name of Provider or Supplier</b>  Planned Parenthood Of Southwest And Central	<b>Street Address, City, State</b>  11500 University Blvd Ste B, Orlando, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the Laboratory Director failed to document the review and evaluation of proficiency testing (PT) for 5 out of 6 testing events reviewed. Findings: Review of American Proficiency Institute (API) proficiency testing results showed that the laboratory director failed to sign the "Proficiency Testing Performance Evaluation" for the Immunology/Immunochemistry PT for the 1st, 2nd, and 3rd events in 2017, and the 1st and 2nd testing events in 2018. During an interview on 11/26/18 at 9:50 AM, the Office Manager acknowledged that the performance evaluation forms were not signed.</p>