

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0901350	(X3) Date Survey Completed 06/14/2022
Name of Provider or Supplier Pediatric Associates - Skylake	Street Address, City, State 1610 Ne Miami Gardens Dr, North Miami Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 06/14/2022 found the PEDIATRIC ASSOCIATES-SKYLAKE clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to score at least 80 percent (%) in Bacteriology for the first proficiency testing (PT) event in 2022. Findings include: Review of Bacteriology American Proficiency Institute (API) PT records revealed a 60 % score in the first event of 2022. During an interview on 06/14 /2022 at 02:25 PM, the area manager confirmed that the laboratory failed to receive a passing score of at least 80 % in Bacteriology in the first event of 2022.</p>