

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 10D0903418	<b>(X3) Date Survey Completed</b> 08/31/2021
<b>Name of Provider or Supplier</b> Personal Care Pediatrics	<b>Street Address, City, State</b> 2964 N Sr 7 Ste 340, Margate, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey conducted on 08/31/2021 found that the PERSONAL CARE PEDIATRICS clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D2020</b>	<p><b>BACTERIOLOGY</b> CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of API (American Proficiency Institute) proficiency testing (PT) records and laboratory director (LD) interview, the laboratory failed to obtain a passing score of 80 % for Bacteriology in 1 out of 6 events reviewed (2019 to 2021). Findings include: Review of API PT records revealed a score of 40 % for Bacteriology in the 1st event of 2021. During an interview on 08/31/2021 at 11:30 AM, the LD confirmed that the laboratory failed the event of reference.</p>