

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0914265	(X3) Date Survey Completed 01/23/2026
Name of Provider or Supplier Pensacola Pediatrics Pa - Tiger Point	Street Address, City, State 1368 Country Club Rd, Gulf Breeze, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Pensacola Pediatrics-Tiger Point PA on 01/23/2026. The laboratory is not in compliance with 42 CFR Part 493, Requirement for Laboratories. The following Condition was cited: D6033 493.1413(b)(8) Condition: Technical Consultant - Moderate Complexity
D6033	<p>TECHNICAL CONSULTANT-MODERATE COMPLEXITY CFR(s): 493.1409</p> <p>The laboratory must have a technical consultant who meets the qualification requirements of 493.1411 of this subpart and provides technical oversight in accordance with 493.1413 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on record review, and interview, the Technical Consultant failed to be responsible for evaluating the competency of all Testing Personnel for 2 of 2 years (2024-2026) reviewed. (See D6046)</p>
D6046	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. The procedures for evaluation of the competency of the staff must include, but are not limited to--</p> <p>This STANDARD is not met as evidenced by: Based on record review, and interview, the Technical Consultant failed to be responsible for evaluating the competency of all Testing Personnel for 2 of 2 years (2024-2026) reviewed. Findings: 1.) Review of the Laboratory Personnel Report</p>

signed by the Laboratory Director on 01/19/2026 showed there were 6 Testing Personnel listed on the form, Testing Personnel (TP) #A, B, C, D, E, with TP# E also listed as a Technical Consultant (TC #B) including the Laboratory Director (LD) who was listed as Clinical Consultant and TC# A. 2.) The Laboratory Procedure manual signed by the Laboratory Director on 09/19/2025 included the Quality Assurance Plan which documents, "At least annually, the laboratory director and/or technical consultant will review the performance of each employee working in the laboratory to assure employee competency." 3.) Review of Annual Evaluations 2024 and 2025 of Testing Personnel Competency for TP# A,B,C,D,E, and LD showed they were signed by TP# A and for TP# A showed signed by TP# B on 05/14/24 and 05/1/25. 4.) Interview with TP# A on 01/23/2026 at 10:55am confirmed that they observed and signed all annual competencies for 2024 and 2025 without the supervision of either the LD, and/or TC# A or B.